

EC Type-Examination Certificate

Number 5628 Issue 3 Extension 2

This is to certify that the product group reference "u-fit lite" comprising the following:

Product reference	Description
u-fit lite	Powder free blue colour nitrile glove. Non-sterile. (Low Dermatitis Potential)

Sizes 6-10	Classification	
	EN 388:2003	Level
	Abrasion resistance	0
	Blade cut resistance	0
	Tear resistance	0
	Puncture resistance	0
	EN 374-2:2003	
	Air Leak	Pass
	Water Leak	Pass
	EN 374-3:2003	
	40% Sodium hydroxide	6
	n-Heptane	0
	5% Ethidium Bromide	6
	37% Formaldehyde	6
	0.1% Phenol	6
	10-13% Sodium Hypochlorite	6
	50% Sulphuric Acid	6
	1.5% Methanol in water	6
	50% Gluteraldehyde	6
	70% Isopropanol	1

Technical Reports

SATRA: CHM0199818/1151(SPC), CHM0199818/1151/MB
 ARDL: PN 81519


has been subject to an EC Type-examination in accordance with Article 10 of the PPE Directive (89/686/EEC) and has been shown to satisfy the relevant provisions of this Directive for the complex category through:

- i Testing to the following standard: **EN374-1:2003 (excluding clause 5.3.2)**
- ii Examination of the relevant technical documentation.

You are therefore licensed to mark the product(s) listed above in accordance with Article 13 of Directive (89/686/EEC) and any relevant amending Directives once you have drawn up an EC declaration of product conformity. Please note that:

1. Full details of the certification and product are contained in the manufacturer's technical file
2. This certificate is only valid if embossed with the text – "SATRA – European Notified Body 0321"
3. This certificate is issued subject to the conditions on the reverse side of this certificate
4. CE marking of production is also reliant on current compliance with Directive 89/686/EEC Article 11

Signed:  (M.J. Cooper)

Signed: 

On behalf of SATRA

Date: 14th September 2013.