



# CB TEST CERTIFICATE

Ref. Certificate No.

CH-6854

## IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

Issued by:	Electrosuisse		
Product:	Light therapy equipment		
Applicant:	Bioptron AG	Gewerbepark Rohrbüel Sihleggstrasse 23 CH-8832 Wollerau	Switzerland
Manufacturer:	Bioptron AG	Gewerbepark Rohrbüel Sihleggstrasse 23 CH-8832 Wollerau	Switzerland
Factory:	Bioptron AG	Gewerbepark Rohrbüel Sihleggstrasse 23 CH-8832 Wollerau	Switzerland
Rating and principal characteristics:	100-240 V~, 50/60 Hz, 0.29-0.12 A Class II, IP20		
Trade mark (if any):	BIOPTRON		
Model/Type reference:	MedAll		
Additional information:	Additional standards tested:	IEC 60601-1-11(ed.1) ICES-001 (Issue 4, 2006-06) FCC CFR47 part 15, edition 2009-10-01, class B	
Sample of product tested to be in conformity with IEC:	CISPR 11(ed.5);am1 60601-1(ed.3) 60601-1-2(ed.3) 61000-3-2(ed.3);am1;am2 61000-3-3(ed.2) 61000-6-3(ed.2);am1	<b>National differences:</b> EU Group Differences; EU Special National Conditions; EU A-Deviations	
Test Report Ref. No.:	11-EL-0070.01 + .02 + .10(EMC)		



This CB Test Certificate is issued by the National Certification Body:

Electrosuisse  
Luppenstrasse 1,  
CH-8320 Fehraltorf

Signed by: Martin Plüss  
2012-10-10

**electrosuisse** >>>