

We can only guarantee the performance and safety of our instruments if the appropriate HEINE bulb and battery powers it. If a non-original copy of a HEINE bulb or battery is used, the safety and performance of the HEINE instrument will be compromised and the manufacturer's guarantee will no longer be valid.

Insist on genuine HEINE replacement bulbs, batteries and parts to ensure performance of your HEINE precision instrument.

HEINE batteries

For optimum performance and cost effectiveness

X-02.99.380 NiCad BETA battery has been replaced by X-07.99.381 or X-02.99.382

X-04.99.623 has been cancelled, use X-04.99.624 or X-04.99.641 in Accubox-II / II L units only.

| | Instrument | 2.5 V | 3.5 V |
|---|---|-------------|-------------|
|  | BETA LI-ION, Lithium-Ion (LI-ION) 3,5V rechargeable battery for: All BETA handles Laryngoscope handles | | X-07.99.381 |
|  | BETA NiMH, Nickel Metal Hydride (NiMH) 3,5V rechargeable battery for: All BETA handles Laryngoscope handles | | X-02.99.382 |
|  | M3Z, Nickel Metal Hydride (NiMH) 3,5V rechargeable battery for: BETA slim handle, Alpha+ handle Small laryngoscope handles | | X-02.99.106 |
|  | M2Z, Nickel Cadmium (NiCad) 2,5V rechargeable battery for: BETA slim handle, Alpha+ handle Small laryngoscope handles | D-01.89.013 | |
|  | S2Z, Nickel Metal Hydride (NiMH) 2,5V rechargeable battery for: Old S2Z handles | X-01.99.333 | |
|  | S3Z, Nickel Metal Hydride (NiMH) 3,5V rechargeable battery for: Old S3Z handles | | X-02.99.314 |
|  | Stand N, Nickel Metal Hydride (NiMH) 2,5V rechargeable battery for: Old Nicatron N, Stand N, handles | X-01.99.433 | |
| | | | 6 V |
|  | HC6V, Nickel Cadmium (NiCad) 6V rechargeable battery for: Old Accubox-II and Accubox-II L | | X-04.99.624 |
| | Nickel Metal Hydride (NiMH) 6V rechargeable battery for: Old Accubox-II and Accubox-II L | | X-04.99.641 |
|  | Lithium-Ion (LI-ION) 6V rechargeable battery for: mPack | | X-07.99.675 |