

CERTIFICATE OF ANALYSIS

ISO 9001-2015 CERTIFIED

Product Name : SODIUM RHODIZONATE AR
Lot No. : G31692109 **Analyzed On** : 03-Sep-2021
Mol. Formula : C6Na2O6 **Mol. Weight** : 214.04
Code No. : 05985 **CAS No.** : 523-21-7
Mfg. Date : Sep-2021 **Exp. Date** : Aug-2026
HAZ. / P.G. : **UN No.** :

Sr.	Tests	Specifications	Results
1	Appearance	Bluish green crystals	Bluish green crystals
2	Assay (by HClO4 titration)	Min 98.5%	98.6%
3	Suitability (for sulphate titration)	Passes test	Passes test
4	Sensitivity for Barium	1:200000	1:200000

CONCLUSION -This above product complies as per the specifications of **LOBA CHEMIE PVT. LTD.**



Shashikant Gaikwad
QC Manager

LOBA CHEMIE PVT. LTD.

Works : Plot No. D-22, MIDC, Tarapur Industrial Area, Tarapur, Boisar, Taluka- Palghar, Dist. Palghar, Pin-401506

Tel: 91.02525-663630/39/34

Regd Office : 107 Wode House Road, Jehanghir Villa, Colaba, Mumbai-400005

Tel: 91.22.6663 6663, Fax: 91.22.22151099

info@lobachemie.com | www.lobachemie.com

LOBA/QC/FM/01

04/10.03.2021