

**PRODUCT**

# Surface Resistance Metre 9V

093-0005



This meter has been designed to measure both surface resistance and resistance to ground. To measure surface material place the meter on the area to be tested and press the test button. For measuring resistance to ground, place the ground lead into the earth leakage socket, connect the crocodile clip to the area to be tested and press the test button.

The meter will test conductive, static dissipative and insulative surfaces using two parallel bars set to give ohms per square readings. Conductive is indicated by green LEDs from  $10^3 - 10^5$ , static dissipative by yellow LEDs from  $10^6 - 10^{11}$  and insulative by red LEDs  $10^{12}+$ .

TECHNICAL SPECIFICATION	TECHNICAL DATA	
Decade change-over	+/- 1/2 Decade on a Logarithmic Scale (3.16 x 10n)	
Resolution	1 Decade	
Accuracy	+/- 1/2 Decade	
Temperature	Operating	5°C to 49°C
	Storage	-15°C to +60°C
Power	9V PP3 Alkaline Battery	
Test voltage	Nominal 9V	
Accuracy	+/- 10%	
Repeatability	+/- 5%	
Weight	6 Ounce	

PRODUCT CODE	DESCRIPTION	SIZE (mm)	QUANTITY
093-0050	Surface Resistance Meter	130 x 70 x 25	Each

To request a quotation or for more information, please call **+44 (0)1473 836200**  
email [info@antistat.co.uk](mailto:info@antistat.co.uk) or visit [www.antistat.co.uk](http://www.antistat.co.uk)

IMPORTANT: This data sheet and its contents (the "Information") belong to Antistat or are licensed to it. No licence is granted for the use of it other than for information purposes in connection with the products to which it relates. No licence of any intellectual property rights is granted. The Information is subject to change without notice and replaces all data sheets previously supplied. The Information supplied is believed to be accurate but Antistat assumes no responsibility for its accuracy or completeness, any error in or omission from it or for any use made of it. Users of this data sheet should check for themselves the Information and the suitability of the products for their purpose and not make any assumptions based on information included or omitted. Liability for loss or damage resulting from any reliance on the Information or use of it (including liability resulting from negligence or where Antistat was aware of the possibility of such loss or damage arising) is excluded. © 2021 Antistat.

## OPERATION

Remove battery panel and insert PP3 Alkaline battery. To check the battery hold down the unit in the air and press the test button, the red LED should show (insulative). Place the meter onto the surface ensuring the probes are touching the area to be tested. and press the Test button. The LED that is illuminated shows the relevant results in Ohms per square.

## CHECKING RESISTANCE TO GROUND

Insert the grounding lead into the earth leakage socket and connect the crocodile clip to the ground connection. Place the meter onto the surface being tested and hold down the test button. The measurement will show the resistance to ground in Ohms rather than Ohms per square.

## NOTES

Calibration should be carried out biannually (please contact us for more information). The meter comes complete with a ground lead. Battery not supplied.

## ARTWORK

Supplied with standard artwork. For a non branded version please contact one of our sales advisors. MOQs may apply.

Surface Resistance Meter packaged in a black corrugated box.

To request a quotation or for more information, please call **+44 (0)1473 836200**  
email [info@antistat.co.uk](mailto:info@antistat.co.uk) or visit [www.antistat.co.uk](http://www.antistat.co.uk)

IMPORTANT: This data sheet and its contents (the "Information") belong to Antistat or are licensed to it. No licence is granted for the use of it other than for information purposes in connection with the products to which it relates. No licence of any intellectual property rights is granted. The Information is subject to change without notice and replaces all data sheets previously supplied. The Information supplied is believed to be accurate but Antistat assumes no responsibility for its accuracy or completeness, any error in or omission from it or for any use made of it. Users of this data sheet should check for themselves the Information and the suitability of the products for their purpose and not make any assumptions based on information included or omitted. Liability for loss or damage resulting from any reliance on the Information or use of it (including liability resulting from negligence or where Antistat was aware of the possibility of such loss or damage arising) is excluded. © 2021 Antistat.