

Page 1 of 1

Analytical Report: AAK00902

Eurofins Sample Number: LV20AB4831-1

Version: 1

Eurofins WKS Labservice GmbH

Industriestr. 1 49849 Wislum

DE

Client Account Number: A00814957PVZ Eurofins Quote Number: N3OPPH20040702

# Eurofins Sample Number LV20AB4831-1

Original Received Date: 14-Sep-2020

**Description:** 1 - Mundschutzmaske Typ II R

Lot Number: 602001

Analysis	Specification	Result	Unit				
Microbial Cleanliness	Microbial Cleanliness (Bioburden)						
Aerobic count		< 0.91	CFU/g				
Aerobic count		< 3.00	CFU/sample				
Yeasts and Moulds		< 0.91	CFU/g				
Yeasts and Moulds		< 3.00	CFU/sample				
Total Bioburden	≤ 30	< 1.82	CFU/g				
Total Bioburden		< 6.00	CFU/sample				
Method: EN 14683:20	019/AC 2019 App D						

Method: EN 14683:2019/AC 2019 App D Analysis Date: 15-Sep-2020 to 22-Sep-2020

#### Sample Compliance Assessment

LV20AB4831-1 meets the requirement(s) for all listed test(s) where specifications were applied.

#### Supplemental Information

WKS-ID: 7116/11030

# Contracted Company: Eurofins Biolab Srl (Vimodrone)

Via Bruno Buozzi, 2, Vimodrone, MI 20055 Italy

InfoFarma@eurofins.com



Page 1 of 1

Analytical Report: AAK00904

Eurofins Sample Number: LV20AB4831-2

Version: 1

Eurofins WKS Labservice GmbH

Industriestr. 1 49849 Wislum

DE

Client Account Number: A00814957PVZ Eurofins Quote Number: N3OPPH20040702

# Eurofins Sample Number LV20AB4831-2

Original Received Date: 14-Sep-2020

**Description:** 2 - Mundschutzmaske Typ II R

Lot Number: 602001

Analysis	Specification	Result	Unit
Microbial Cleanlines	ss (Bioburden)		
Aerobic count		< 0.91	CFU/g
Aerobic count		< 3.00	CFU/sample
Yeasts and Moulds		<0.91	CFU/g
Yeasts and Moulds		< 3.00	CFU/sample
Total Bioburden	≤ 30	< 1.82	CFU/g
Total Bioburden		<6.00	CFU/sample
Method: EN 14683:	2019/AC 2019 App D		

Method: EN 14683:2019/AC 2019 App D Analysis Date: 15-Sep-2020 to 22-Sep-2020

#### Sample Compliance Assessment

LV20AB4831-2 meets the requirement(s) for all listed test(s) where specifications were applied.

#### Supplemental Information

WKS-ID: 7116/11030

# Contracted Company: Eurofins Biolab Srl (Vimodrone)

Via Bruno Buozzi, 2, Vimodrone, MI 20055 Italy

InfoFarma@eurofins.com



Page 1 of 1

Analytical Report: AAK00901

Eurofins Sample Number: LV20AB4831-3

Version: 1

Eurofins WKS Labservice GmbH

Industriestr. 1 49849 Wislum

DE

Client Account Number: A00814957PVZ Eurofins Quote Number: N3OPPH20040702

# Eurofins Sample Number LV20AB4831-3

Original Received Date: 14-Sep-2020

**Description:** 3 - Mundschutzmaske Typ II R

Lot Number: 602001

Analysis	Specification	Result	Unit
Microbial Cleanliness	(Bioburden)		
Aerobic count		< 0.91	CFU/g
Aerobic count		<3.00	CFU/sample
Yeasts and Moulds		< 0.91	CFU/g
Yeasts and Moulds		< 3.00	CFU/sample
Total Bioburden	≤ 30	< 1.82	CFU/g
Total Bioburden		< 6.00	CFU/sample
Method: EN 14683:20	019/AC 2019 App D		·

Method: EN 14683:2019/AC 2019 App D Analysis Date: 15-Sep-2020 to 22-Sep-2020

#### Sample Compliance Assessment

LV20AB4831-3 meets the requirement(s) for all listed test(s) where specifications were applied.

#### Supplemental Information

WKS-ID: 7116/11030

# Contracted Company: Eurofins Biolab Srl (Vimodrone)

Via Bruno Buozzi, 2, Vimodrone, MI 20055 Italy

InfoFarma@eurofins.com



Page 1 of 1

Analytical Report: AAK00903

Eurofins Sample Number: LV20AB4831-4

Version: 1

Eurofins WKS Labservice GmbH

Industriestr. 1 49849 Wislum

DE

Client Account Number: A00814957PVZ Eurofins Quote Number: N3OPPH20040702

# Eurofins Sample Number LV20AB4831-4

Original Received Date: 14-Sep-2020

**Description:** 4 - Mundschutzmaske Typ II R

Lot Number: 602001

Analysis	Specification	Result	Unit
Microbial Cleanlines	s (Bioburden)		
Aerobic count		< 0.89	CFU/g
Aerobic count		< 3.00	CFU/sample
Yeasts and Moulds		< 0.89	CFU/g
Yeasts and Moulds		< 3.00	CFU/sample
Total Bioburden	≤ 30	< 1.78	CFU/g
Total Bioburden		< 6.00	CFU/sample
Method: EN 14683:2	2019/AC 2019 App D		

Method: EN 14683:2019/AC 2019 App D Analysis Date: 15-Sep-2020 to 22-Sep-2020

#### Sample Compliance Assessment

LV20AB4831-4 meets the requirement(s) for all listed test(s) where specifications were applied.

#### Supplemental Information

WKS-ID: 7116/11030

# Contracted Company: Eurofins Biolab Srl (Vimodrone)

Via Bruno Buozzi, 2, Vimodrone, MI 20055 Italy

InfoFarma@eurofins.com



Page 1 of 1

Analytical Report: AAK00900

Eurofins Sample Number: LV20AB4831-5

Version: 1

Eurofins WKS Labservice GmbH

Industriestr. 1 49849 Wislum

DE

Client Account Number: A00814957PVZ Eurofins Quote Number: N3OPPH20040702

# Eurofins Sample Number LV20AB4831-5

Original Received Date: 14-Sep-2020

**Description:** 5 - Mundschutzmaske Typ II R

Lot Number: 602001

Analysis	Specification	Result	Unit
Microbial Cleanlines	s (Bioburden)		
Aerobic count		< 0.92	CFU/g
Aerobic count		< 3.00	CFU/sample
Yeasts and Moulds		< 0.92	CFU/g
Yeasts and Moulds		< 3.00	CFU/sample
Total Bioburden	≤ 30	< 1.84	CFU/g
Total Bioburden		< 6.00	CFU/sample
Method: FN 14683:2	019/AC 2019 App D	·	·

Method: EN 14683:2019/AC 2019 App D Analysis Date: 15-Sep-2020 to 22-Sep-2020

#### Sample Compliance Assessment

LV20AB4831-5 meets the requirement(s) for all listed test(s) where specifications were applied.

#### Supplemental Information

WKS-ID: 7116/11030

# Contracted Company: Eurofins Biolab Srl (Vimodrone)

Via Bruno Buozzi, 2, Vimodrone, MI 20055 Italy

InfoFarma@eurofins.com







#### LAB Nº 1827 L

TEST REPORT	Refer to Analytical Report N	lumber						
	Eurofins WKS Labservice GmbH							
Sponsor	Industriestr. 1							
o. oook	49849 Wislum	19849 Wislum						
	GERMANY							
TEST METHOD	Bacterial Filtration Efficiency	/ (BFE) – EN 14683:2019+AC:2019	Арр В					
TEST ITEM - INFORMATION FRO	OM THE SPONSOR							
PRODUCT NAME	Mundschutzmaske Typ II R							
MATRIX OF THE PRODUCT	Face Mask							
Ватсн	602001	CODE	WKS-ID: 7116/	11030				
EUROFINS COSMETICS & PERS	SONAL CARE ITALY IDENTIFICATION	ON						
MATERIAL ITEM ALIQUOT	LV20AB4826-1							
PARCEL REGISTRATION N.	IP-LV-2020258-ADA	RECEIVING DATE	14 Sep 2020					
Analysis Starting Date	15 Sep 2020	ANALYSIS ENDING DATE	17 Sep 2020					
EXPERIMENTAL CONDITIONS	Size of the area tested: 49 of Flow rate during testing: 28, Inner side of the mask to the	3 l/min	Table Canada					
	Size of the area tested: 49 of Flow rate during testing: 28,	cm² 3 l/min	1					
PHOTO OF THE TEST ITEM  RESULTS	Size of the area tested: 49 of Flow rate during testing: 28,	RES 98 98 98	SULT UNIT ,40 % ,72 % ,83 % ,87 %					

This test report may not be reproduced in part unless expressly approved in writing by Eurofins Cosmetics & Personal Care Italy Srl. The test results relate only to the tested items. Sampling, except specific indication on test report, is always intended to be made by the Sponsor. Information on the test item provided by the Sponsor are under Sponsor responsibility.

Tel: +39-022507151 - Fax: +39-0225071599 - E-mail: : InfoCosme@eurofins.com







LAB N° 1827 L

Page: 2 of 2

#### Addendum N.1

Started on: 15/09/2020 LV20AB4826 Batch:

Sample description: Mundschutzmaske Typ II R

Lot Number: 602001

#### **Negative Control Plate Counts**

	Stage 1	Stage 2	Stage 3*	Stage 4*	Stage 5*	Stage 6*	Mean
Negative Control (CFU)	0	0	0	0	0	0	0

<sup>\*</sup>number of colonies adjusted with positive-hole correction table

#### Positive Controls Plate Counts

	Stage 1	Stage 2	Stage 3*	Stage 4*	Stage 5*	Stage 6*	T-4-LOFU
Size of particle (µm)	7,00	4,70	3,30	2,10	1,10	0,65	Total CFU
Positive Control N.1 (CFU)	137	341	902	559	426	287	2652
Positive Control N.2 (CFU)	171	327	1160	527	475	171	2831

number of colonies adjusted with positive-hole correction table

Mean of the total plate counts of the two positive controls (CFU):

2742

#### Mean Particle Size (MPS)

	MPS
Positive Control N.1 (μm)	2,78
Positive Control N.2 (µm)	2,93
Mean (µm)	2,86

#### Test specimens Plate Counts

	Stage 1	Stage 2	Stage 3*	Stage 4*	Stage 5*	Stage 6*	Total CFU
LV20AB4826-1 - Aliquot 1	0	0	0	1	19	24	44
LV20AB4826-1 - Aliquot 2	0	0	0	1	12	22	35
LV20AB4826-1 - Aliquot 3	0	0	0	1	6	25	32
LV20AB4826-1 - Aliquot 4	0	0	0	0	9	22	31
LV20AB4826-1 - Aliquot 5	0	0	0	2	8	15	25

<sup>&</sup>quot;number of colonies adjusted with positive-hole correction table

# Test specimens Bacterial Filtration Efficiency (BFE)

	BFE (%)
LV20AB4826-1 - Aliquot 1	98,40
LV20AB4826-1 - Aliquot 2	98,72
LV20AB4826-1 - Aliquot 3	98,83
LV20AB4826-1 - Aliquot 4	98,87
LV20AB4826-1 - Aliquot 5	99,09

Eurofins Cosmetics & Personal Care Italy Srl - via B.Buozzi 2, Vimodrone (Milano), Italy -P.IVA / VAT Number: 05533561006

Tel: +39-022507151 – Fax: +39-0225071599 – E-mail: : <u>InfoCosme@eurofins.com</u>



Page: 1 / 6

Analytical Report Nr. Sample code Nr. Date AR-20-YL-005606-01 560-2020-00006099

18/09/2020

# ANALYTICAL REPORT

#### **Client Information**

Eurofins WKS Labservice GmbH Industriestraße 1 Wilsum GERMANY 0049 5945 89220 labor-wilsum@eurofins-wks.de

For the attention of David Kerperin

**Sample Information** 

Order Code: EUAA70-0008165

Reception Date:14-Sep-2020Analysis Starting Date:14-Sep-2020Analysis Ending Date:18-Sep-2020

Sample described as: Masks

Information provided by the customer:

Client Reference: 7116/11030

Sample Description:

Customer requirements: EN 14683:2019+AC:2019 TYPE IIR

**Purchase Order Number:** 

Decision rule Shared risk - Simple Batch Not provided

acceptance.







Analytical Report Nr.
Sample code Nr.
Date

AR-20-YL-005606-01 560-2020-00006099 18/09/2020

# **SAMPLE PICTURE**





Page: 3 / 6

Analytical Report Nr.
Sample code Nr.
Date

AR-20-YL-005606-01 560-2020-00006099 18/09/2020

# **CONCLUSION:**

TEST PROPERTY	PASS	FAIL	REMARKS
Breathability (Differential Pressure)	1		
EN 14683:2019+AC:2019 Annex C	1		
	1		
A - Mask	X		
	! !		
Resistance against penetration by synthetic blood	1		
ISO 22609:2004	1 1 1		
	1		
A - Mask	X		
	1 1		

**Remark:** Test has been performed as per application request





Page: 4 / 6

Analytical Report Nr. Sample code Nr. Date AR-20-YL-005606-01 560-2020-00006099 18/09/2020

### COMPONENT LIST:

COMPONENT ID	COMPONENT NAME	MATERIAL DESCRIPTION	COLOR	REMARKS
CUST 01	A - Mask	Mask	Blue	



Page: 5 / 6

Analytical Report Nr. Sample code Nr.

**Date** 

AR-20-YL-005606-01 560-2020-00006099

18/09/2020

MASKS TESTING CAS No. RESULTS UNC. LOQ GUIDELINES

Analyses on: A - Mask

Resistance against penetration by synthetic blood

Analysis Ending Date: 16/09/2020

Analysis Ending Date: 18/09/2020

ISO 22609:2004

Result: 31 of 32 pass at 16 KPa - 29 of 32 samples 🗸 Pass

pass at 16 kPa

Complete test data reported at Annex.

AQL information according to ISO 22609:2004:
A single sampling plan providing an AQL of 4,0 % requires 32 test specimens.
An AQL of 4.0 % is met for a single sampling plan when 29 or more specimens show pass results.
AQL= Acceptable Quality Limit.

**Breathability (Differential Pressure)** 

EN 14683:2019+AC:2019 Annex C

Differential pressure 43.9 Pa/cm² (± 1.5) Pa/cm² - <60 Pa/cm² ✓ Pass

Complete test data reported at Annex.



Page: 6 / 6

Analytical Report Nr.
Sample code Nr.
Date

AR-20-YL-005606-01 560-2020-00006099 18/09/2020

#### Signed for and on behalf of Eurofins Textile Testing Spain:



Report electronically validated by

Maria Jesus Martinez Puig Chemical Lab manager

#### **EXPLANATORY NOTE**

- Test not covered by ENAC accreditation scope
- Test is subcontracted within Eurofins group and is accredited
- Test is subcontracted within Eurofins group and is not accredited
- Test is subcontracted outside Eurofins group and is accredited
- □ Test is subcontracted outside Eurofins group and is not accredited N/A = Not Applicable

Eurofins General Sales Terms and Conditions applied.

Results obtained refer only to samples, products or material received in Laboratory, as described in point related to sample description, and tested in conditions shown in present report.

Test uncertainties not reported are at customer disposal.

If you happen to have any comments, please do it by sending email to textile\_spain@eurofins.com and referring to this report number.

Reproduction of this document is only valid if it is done completely and under the written permission. Any printed version of this document will be considered as a copy.

**End Of Report** 





#### **DETERMINATION RESISTANCE AGAINST PENETRATION BY SYNTETIC BLOOD**

Test Method: ISO 22609:2004; Targeting-plate test method

Number of test specimens: 32

Sample size: Circular, diameter 5,58 cm

Sample area tested: 24,5 cm<sup>2</sup>

Pressure: 16 kPa (120,0 mm Hg)

Stream velocity of synthetic blood: 550±10 cm/s

Distance of the face mask target area surface from the tip of the cannula: 30,5 cm

Angle of the pneumatic valve with respect to the face mask target area: 90°

Technique used to enhance visual detection of synthetic blood: Hydrophilic cotton

Conditioning: At least 4 hours. Ta between 16,7°C and 26 °C. HR% between 82,8 and 88 % Hr

Environmental test conditions 24,1 °C; 78,7 % Hr

Pre-treatment: None

	7	R	esults				
Specimen			Pass	Fail			
1			X				
2			X				
3			X				
4			X				
5			X				
6			X				
7			X				
8	U. T. C.		X				
9			X				
10			X				
11			7		X		
12			X		4		
13			X		7		
14			X		7		
15			X				
16			X				
17			X				
18			X				
19			X				
20			X				
21			X				
22			Х				
23		V	X	7			
24		7	X	7			
25	1	7	X	and the second			
26			X				
27			X				
28			X				
29			X				
30			X				
31			X				
32			X				

Conclusion	PASS
	I FASS



# METHOD FOR DETERMINATION OF BREATHABILITY (DIFFERENTIAL PRESSURE)

Test Method: EN 14683: 2019+AC: 2019 Annex C

Number of test specimens: 5


Number of test per specimen: 5

Sample area tested: Circular, diameter 2,5 cm

Tested area of the test sample: 4,9 cm<sup>2</sup>

Flow rate during testing: 8±0,25 l/min

General location of measurement areas: Representative of the overall surface.



### Results

	Units (Pa)					ΔΡ	
Specimen	Position	Position	Position	Position	Position	Mean value	(Pa/cm <sup>2</sup> )
	1	2	3	4	5	(Pa)	(Pa/CIII <sup>2</sup> )
1	219	204	229	205	197	211	43,0
2	218	227	224	229	215	223	45,4
3	221	215	203	211	205	211	43,1
4	231	225	245	224	198	225	45,8
5	203	207	206	214	204	207	42,2
			7			Mean Value	43,9
					Uncertainty	± 1,5	

# Observation:

For thick and rigid masks the test method may not be suitable as a proper seal cannot be maintained in the sample holder.